



Evos Turn OFF

Check the booking calendar. Are you the last user?

α

Close M7000 software




β

1. remove your sample from the microscope
2. clean the objective lens using lens cleaning tissue

γ

Turn off Evos microscope



δ

Turn off incubator

