

The chamber should be left at 37oC as default, unless you need to carry out your imaging at a different temperature.

In such circumstances, the user requiring the lower temperature should write a comment on the booking system to notify any subsequent user and set the temperature back to 37oC at the end of their session.

The chamber requires some time to change temperature, therefore an hour should be left so that the system can equilibrate to the new temperature – both to cool down to room temperature and to warm up to 37oC. If this is the case then simply add a note to the booking system and no charge will be made for these 2 hours extra.