



Zeiss Fluorescence Turn OFF

Check the booking calendar. Are you the last user?

α

Close Metamorph software
Log off computer



β

1. remove your sample from the microscope
2. clean the objective lens using lens cleaning tissue

γ

Turn off controller box at the mains socket

